

**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME**

**CB TEST CERTIFICATE**

Product	Power Supply
Name and address of the applicant	NXT POWER LLC 7051 W Wilson Ave Harwood Heights, IL, 60706 USA
Name and address of the manufacturer	NXT POWER LLC 7051 W Wilson Ave Harwood Heights, IL, 60706 USA
Name and address of the factory <i>Note: When more than one factory, please report on page 2</i>	NXT POWER LLC 7051 W Wilson Ave Harwood Heights, IL, 60706 USA  <input type="checkbox"/> <a href="#">Additional Information on page 2</a>
Ratings and principal characteristics	See Page 2
Trademark / Brand (if any)	NXT POWER LLC
Type of Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	NPT80-XXX-i, NPT240-XXX-i, NPT480-XXX-i, NPT780-XXX-i, NPT1000-XXX-i, NPT1200-XXX-i, NPT1500-XXX-i, NPT2000-XXX-i
Additional information (if necessary may also be reported on page 2)	The report was revised to include corrections. <input checked="" type="checkbox"/> <a href="#">Additional Information on page 2</a>
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005
As shown in the Test Report Ref. No. which forms part of this Certificate	E515677-D1001-1/A0/C1-CB issued on 2020-09-21

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

**Date:** 2020-09-21  
Original Issue Date: 2020-09-17

**Signature:**

Jolanta M. Wroblewska

## Ratings:

NPT80-XXX-i: Input 200-240 Vac, 0.40A, 50/60 Hz; Output 200-240 Vac, 0.35A, 50/60 Hz.  
NPT240-XXX-i: Input 200-240 Vac, 1.15A, 50/60 Hz; Output 200-240 Vac, 1.00A, 50/60 Hz.  
NPT480-XXX-i: Input 200-240 Vac, 2.20A, 50/60 Hz; Output 200-240 Vac, 2.00A, 50/60 Hz.  
NPT780-XXX-i: Input 200-240 Vac, 3.58A, 50/60 Hz; Output 200-240 Vac, 3.25A, 50/60 Hz.  
NPT1000-XXX-i: Input 200-240 Vac, 4.51A, 50/60 Hz; Output 200-240 Vac, 4.10A, 50/60 Hz.  
NPT1200-XXX-i: Input 200-240 Vac, 5.94A, 50/60 Hz; Output 200-240 Vac, 5.40A, 50/60 Hz.  
NPT1500-XXX-i: Input 200-240 Vac, 7.15A, 50/60 Hz; Output 200-240 Vac, 6.50A, 50/60 Hz.  
NPT2000-XXX-i: Input 200-240 Vac, 9.35A, 50/60 Hz; Output 200-240 Vac, 8.50A, 50/60 Hz.

## Additional Information:

National Differences are specified in the CB Test Report.

## Model Differences

## XXX:

M= Medical unit  
MH- Medical Hard wire unit.  
MW= Medical with wall mount unit  
MHW- Medical Hard wire with wall mount unit.  
MF= Medical with no neutral to ground bond unit  
MFW= Medical no neutral to ground bond with wall mount unit  
MFHW= Medical Hard wire no neutral to ground bond with wall mount unit

The Risk Management requirements of the standard were not addressed.

The original report was modified to include the following changes/additions:  
To correct typographical errors in model information. See Test Report.

## Additional information (if necessary)



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN


UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/incbnames](http://www.ul.com/incbnames)

Date: 2020-09-21

Original Issue Date: 2020-09-17

Signature:



Jolanta M. Wroblewska