



Ref. Certif. No.

US-36616-A2-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Power Supply

Name and address of the applicant

NXT Power, LLC
1070 S Northpoint Blvd (Unit D) Waukegan, IL 60085
United States

Name and address of the manufacturer

NXT Power, LLC
1070 S Northpoint Blvd (Unit D) Waukegan, IL 60085
United States

Name and address of the factory

NXT Power, LLC
1070 S Northpoint Blvd (Unit D) Waukegan, IL 60085
United States

Note: When more than one factory, please report on page 2

Additional Information on page 2

Ratings and principal characteristics

Additional Information on page 2

Trademark / Brand (if any)

NXT POWER LLC

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

NPT80-XXX-i, NPT240-XXX-i, NPT480-XXX-i, NPT780-XXX-i,
NPT1000-XXX-i, NPT1200-XXX-i, NPT1500-XXX-i, NPT2000-XXX-i.
 Additional Information on page 2

Additional information (if necessary may also be reported on page 2)

The report was revised to include administrative modifications.
The risk management requirements of the standard were not addressed
National Differences: CA, US
 Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012

As shown in the Test Report Ref. No. which forms part of this Certificate

E515677-D1001-1/A2/C1-CB issued on 2024-01-17

This CB Test Certificate is issued by the National Certification Body



- UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2024-01-19

Signature:

Original Issue Date: 2020-09-17

Jolanta M. Wroblewska



Ref. Certif. No.

US-36616-A2-UL

Additional Ratings:

NPT80-XXX-i: Input 200-240 Vac, 0.40A, 50/60 Hz, Output 200-240 Vac, 0.35A, 50/60 Hz.
 NPT240-XXX-i: Input 200-240 Vac, 1.15A, 50/60 Hz, Output 200-240 Vac, 1.00A, 50/60 Hz.
 NPT480-XXX-i: Input 200-240 Vac, 2.20A, 50/60 Hz, Output 200-240 Vac, 2.00A, 50/60 Hz.
 NPT780-XXX-i: Input 200-240 Vac, 3.58A, 50/60 Hz, Output 200-240 Vac, 3.25A, 50/60 Hz.
 NPT1000-XXX-i: Input 200-240 Vac, 4.51A, 50/60 Hz, Output 200-240 Vac, 4.10A, 50/60 Hz.
 NPT1200-XXX-i: Input 200-240 Vac, 5.94A, 50/60 Hz, Output 200-240 Vac, 5.40A, 50/60 Hz.
 NPT1500-XXX-i: Input 200-240 Vac, 7.15A, 50/60 Hz, Output 200-240 Vac, 6.50A, 50/60 Hz.
 NPT2000-XXX-i: Input 200-240 Vac, 9.35A, 50/60 Hz, Output 200-240 Vac, 8.50A, 50/60 Hz.

Model Differences:

XXX:
 M= Medical unit
 MH- Medical Hard wire unit.
 MW= Medical with wall mount unit
 MHW- Medical Hard wire with wall mount unit.
 MF= Medical with no neutral to ground bond unit,
 MFW= Medical no neutral to ground bond with wall mount unit
 MFHW= Medical Hard wire no neutral to ground bond with wall mount unit

Summary of Modifications:

Address for applicant, manufacturer and factory was updated.
 Updated UL logo and manual.

Additional information (if necessary)



- UL Solutions (US), 333 Pflingsten Rd IL 60062, Northbrook, USA
- UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2024-01-19
 Original Issue Date: 2020-09-17

Signature:

Jolanta M. Wroblewska